

Information to help you travel more safely in motor vehicles while seated in your wheelchair





www.travelsafer.org www.rercwts.org



When traveling in a motor vehicle, it is generally safest for wheelchair users to transfer to a vehicle seat and use the vehicle seatbelt system or a child safety seat that complies with federal safety standards. The wheelchair should then be stored and secured in the vehicle.


If transferring is not feasible, it is very important to secure the wheelchair to the vehicle facing forward and to use crash-tested seatbelts for the wheelchair-seated rider.

START WITH THE RIGHT EQUIPMENT

▼ It is best if you have a wheelchair that has been designed and tested for use as a seat in motor vehicles, often referred to as a WC19 wheelchair or a transit wheelchair. These wheelchairs comply with ANSI/ RESNA WC19, a voluntary standard developed by safety and rehabilitation experts. Wheelchairs that meet the design and performance requirements of this standard will be labeled to show that they comply with WC19.

Most importantly, a WC19 wheelchair has four, crash-tested securement points where tiedown straps and hooks can be easily attached. These points are clearly marked with a hook symbol.

▼ If a WC19 wheelchair is not available, the next best choice is a wheelchair with an accessible metal frame where tiedown straps and hooks can be attached at frame junctions.



The Wheelchair Tiedown and Occupant Restraint System (WTORS)



▼ It is important to use a complete WTORS to secure the wheelchair and provide the wheelchair occupant with a properly designed and tested lap/ shoulder belt system.

Always use a WTORS that has been crash tested and labeled as complying with ANSI/RESNA WC18 or SAE J2249, a voluntary standard developed by safety and rehabilitation experts. The most common type of wheelchair tiedown uses four straps to secure the wheelchair to the vehicle. Although it requires someone other than the wheelchair rider to secure and release the wheelchair, this tiedown can secure a wide range of WC19 and non-WC19 wheelchairs.

▼ To protect the rider during a crash or sudden braking, and to minimize the likelihood of injury caused by contact with the vehicle, a seatbelt system with both lap and shoulder belts must be used.



Four-Point Tiedowns

Always position the wheelchair and rider facing forward in the vehicle.


When securing a WC19 wheelchair, attach the four tiedown straps to the securement points provided on the wheelchair. Tighten the straps to remove all slack.

If you do not have a WC19 wheelchair, it is best to attach the tiedown straps to welded junctions of the wheelchair frame or to other structural areas where the frame is fastened together with hardened steel bolts - indicated by six raised lines or bumps on the bolt head.



Do not attach tiedowns to adjustable, moving, or removable parts of the wheelchair such as armrests, footrests, and wheels.





When securing non-WC19 wheelchairs,


choose structural securement points as close to the seat surface as possible to provide greater wheelchair stability during travel. It is best if the rear securement points are high enough to result in angles of the rear tiedown straps between 30 and 45 degrees to the horizontal.

If you have a non-WC19 wheelchair with a tilt seat, make sure to attach both the front and rear straps to either the seat frame or to the base frame. Mixing wheelchair securement points between the seat and base can result in the tiedown straps becoming slack if the angle of the seat changes during a crash.

It is best if floor anchor points for rear tiedown straps are located directly behind the rear securement points on the wheelchair. If possible, the front tiedown straps should anchor to the floor at points that are spaced wider than the wheelchair to increase lateral stability during travel.

Other Methods of Wheelchair Securement

In addition to securing wheelchairs using a four-point tiedown, wheelchairs can also be secured using a docking tiedown device. This method is primarily used in private vehicles since it requires the addition of adaptor hardware to the wheelchair frame that will engage with the docking tiedown device in the vehicle. Docking securement devices allow the wheelchair rider to secure and release the wheelchair without assistance.



If you plan to secure your wheelchair with a docking tiedown device, you should check with the WTORS or wheelchair manufacturer to ensure that your wheelchair model has been successfully crash tested with their system.


Clamp-type securement devices are not recommended since they do not provide effective wheelchair securement in frontal crash testing.



PROTECT THE WHEELCHAIR RIDER

In addition to securing the wheelchair, it is very important to provide effective restraint for the wheelchair user with a crash-tested lap and shoulder belt or with a child restraint harness. Postural support belts attached to the wheelchair are not strong enough to withstand crash forces and are usually not positioned correctly to restrain the occupant safely in a crash.

▼ The lap belt should be placed low across the front of the pelvis near the upper thighs, not high over the abdomen. When possible, the lap belt should be angled between 45 and 75 degrees to the horizontal when viewed from the side. Some wheelchair features, like armrests, can interfere with good belt fit. To avoid placing the lap belt over the armrest and to keep the lap belt low on the pelvis, it may be necessary to insert the belt between the armrest and the seatback, or through openings between the backrest and seat.





A diagonal shoulder belt should

cross the middle of the shoulder and the center of the chest, and should connect to the lap belt near the hip of the wheelchair rider. The upper shoulder-belt anchor point or D-ring guide should be anchored above and behind the top of the occupant's shoulder, so that the belt is in good contact with the shoulder and chest while traveling.

▼ Newer WC19 wheelchairs offer the option of a crash-tested lap belt that is anchored to the wheelchair frame. If the wheelchair has an onboard crash-tested lapbelt, complete the belt system by attaching the lower end of a shoulder belt to the lap belt. Crashtested wheelchair-anchored lap belts will be labeled to indicate that they comply with WC19.

Other Important Points

- Read and follow all manufacturers' instructions.
- It is best to ride with the wheelchair backrest positioned at an angle of 30 degrees or less to the vertical. If a greater recline angle is needed, the shoulder belt anchor point should be moved rearward along the vehicle sidewall so the belt maintains contact with the rider's shoulder and chest.
- Maximize the clear space around the rider to reduce the possibility of contact with vehicle components and other passengers in a crash. Cover rigid vehicle components that are close to the rider with dense padding.
- Check wheelchair and WTORS equipment regularly and replace worn components. If involved in a vehicle crash, check with the manufacturer to determine if the equipment needs to be replaced. Keep WTORS anchorage track free of debris.
- If possible, remove hard trays and secure them in the vehicle to reduce the chance of rider injury from contact with the tray. Consider the use of foam trays instead of rigid trays during transit. If it is not possible to remove a hard tray, place dense padding between the rider and the edge of the tray and make sure that the tray is securely attached to the wheelchair so it will not break loose and cause injury to other occupants in a crash.
- A properly positioned headrest may help protect the neck in a rear impact.
- If it is necessary to use a head and neck support during travel, choose a soft, light, neck collar because stiff collars and head straps are more likely to cause neck injury in a crash. The soft collar should not be attached to the seating system.
- Secure medical and other equipment to the wheelchair or vehicle to prevent it from breaking loose and causing injuries in a crash.
- Seating systems can be crashed tested to ANSI/RESNA WC20 and then used with a WC19-compliant frame to create a crashworthy wheelchair.



Institute

www.umtri.umich.edu

RESOURCES

Organizations

Rehabilitation Engineering Research Center on Wheelchair Transportation Safety www.rercwts.org University of Michigan Transportation Research

University of Pittsburgh www.wheelchairnet.org

RESNA Rehabilitation Engineering and Assistive Technology Society of North America www.resna.org

Wheelchair and Seating Manufacturers (Ask for Products that have been Successfully Tested to WC19 and/or WC20)

Bergeron Health Care www.specialtomato.com; 866-529-8407 Colours N Motion www.colourswheelchair.com; 800-892-8998 Convaid www.convaid.com; 888-266-8243 Freedom Designs www.freedomdesigns.com; 800-331-8551 Gillette Children's Specialty Healthcare www.gillettechildrens.org; 800-719-4040 Gunnell www.gunnell-inc.com; 800-551-0055 Hoveround www.hoveround.com; 800-542-7236 Innovative Products www.mobility4kids.com; 800-950-5185 Invacare www.invacare.com; 800-333-6900 Kids Up www.kidsupco.com; 877-454-3787 Ki Mobility www.kimobility.com; 800-981-1540

Metalcraft Industries www.metalcraft-industries.com; 888-399-3232 Motion Concepts www.motionconcepts.com; 888-433-6818

Mulholland Postioning Systems www.mulhollandinc.com; 800-543-4769

Otto Bock www.ottobock.com; 800-328-4058

Performance Health Products www.v-trak.com; 866-632-1755

Permobil www.permobil.com; 800-736-0925

Pride Mobility www.pridemobility.com; 800-800-8586

Product Design Group www.pdgmobility.com; 888-858-4422

Sammons Preston www.sammonspreston.com; 800-323-5547

Stealth Products www.stealthproducts.com; 800-965-9229

Sunrise Medical www.sunrisemedicalonline.com; 800-333-4000 Tilite

www.tilite.com; 800-545-2266

Varalite www.varalite.com; 800-827-4548

Wheelchair Tiedown and Occupant Restraint Manufacturers (Ask for Products that Comply with WC18 or SAE J2249)

EZ-Lock www.ezlock.net; 225-214-4620

New Haven www.safehaven-usa.com; 800-421-8700

Orthosafe www.orthosafe.com; 609-587-9444 Q'Straint www.qstraint.com; 800-987-9987 SureLok www.sure-lok.com; 866-787-3565



GLOSSARY OF TERMS

Anchor point: The location on a vehicle, wheelchair, or wheelchair tiedown where a belt-restraint or wheelchair-tiedown anchorage is attached.

ANSI/RESNA WC18 (officially SECTION 18 RESNA WC-4:2010): A voluntary standard that specifies design and performance requirements for WTORS. NOTE: ISO 10542 is an international standard that is comparable with WC18 and SAE J2249.

ANSI/RESNA WC19 (officially SECTION 19 RESNA WC-4:2010): A voluntary standard for wheelchairs designed for use as a seat when traveling in a motor vehicle. NOTE: ISO 7176-19 is an international wheelchair standard that is comparable with WC19.

ANSI/RESNA WC20 (officially SECTION 20 RESNA WC-4:2010): A voluntary standard for wheelchair seating systems designed or used as part of a wheelchair when traveling in a motor vehicle. NOTE: ISO 16840-4 is an international wheelchair standard that is comparable with WC20.

SAE Recommended Practice J2249: A Society of Automotive Engineers Recommended Practice that specifies design and performance requirements for WTORS. NOTE: WC18 is an enhanced version of this standard and ISO 10542 is a similar international standard.

Belt: A length of energy-absorbing webbing material used in occupant restraint systems.

Docking tiedown: A method for securing wheelchairs where portions of the wheelchair frame, or add-on components fastened to the wheelchair frame, engage with a securement device anchored to the vehicle.

Four-point strap-type tiedown: A method for securing a wheelchair where four straps are attached to the wheelchair at four separate securement points and attached to the vehicle at four separate anchor points.

Occupant restraint: A system or device designed to restrain a motor vehicle occupant in a crash by keeping the occupant in the vehicle seat and minimizing contact with the vehicle interior, other occupants, or objects outside the vehicle.

Postural support: A padded component and/or belt used to help maintain a person in a desired position during normal wheelchair use. In general postural supports are <u>not</u> designed to provide effective occupant restraint in a motor vehicle crash.

Securement points: Specific structural points on the wheelchair base or seat frame that are designed for attachment of wheelchair tiedowns.

Strap: A length of webbing material used in wheelchair tiedown systems.

WC19 wheelchair: A crash-tested wheelchair with four clearly identified securement points that meets the design and performance requirements of ANSI/RESNA WC19 and is sometimes called a transit wheelchair.

WC20 seating system: A crash-tested seating system and its attachment hardware that meets the design and performance requirements of ANSI/RESNA WC20 and is used with a WC19 compliant frame to create a crashworthy wheelchair.

Wheelchair tiedown and occupant-restraint system (WTORS): A complete system for use by wheelchairseated occupants comprised of a system or device for securing the wheelchair and a belt-type restraint system for limiting occupant movement in a motor vehicle crash.

Rehabilitation Engineering Research Center on Wheelchair Transportation Safety

University of Michigan Health System

University of Michigan Transportation Research Institute

Initially funded through a grant from the FRIENDS of the University of Michigan Hospitals

2010

